




Refer to User Manual for comprehensive information:
<https://www.mttS-asia.com/impala-ventilator/>

Set up




No Name


- | No | Name |
|----|--|
| 1 | Main unit |
| 2 | 12V DC adapter |
| 3 | Oxygen hose |
| 4 | Straight connector (ID=15mm OD=22mm) |
| 5 | Exhalation valve |
| 6 | Inspiratory tube (ID=22mm L=120cm) |
| 7 | Exhalatory tube (ID=22mm L=60cm) |
| 8 | Straight connector (OD=22mm OD=22mm) |
| 9 | Wye connector (OD=22mm) |
| 10 | Flow sensor |
| 11 | HMEF (Humidity Moisture Exchange Filter) |

Operation


1 Press and hold power button to turn device ON/OFF.




2 Press OK to continue with the last patient or change patient profile to start with a new patient





3 Follow instructions on the screen to calibrate oxygen sensor, flow sensor and to perform tightness test


 Skipping these steps may result in measurement inaccuracy





4 Choose ventilation mode



5 Set appropriate parameters



6 Set alarm limits



7 Start / stop ventilation

